

Technical Data

Nominal voltage: AC220V

Nominal frequency: 60HZ

Main fuses: 10A

Nominal power: 1000 VA

Dry weight, unpacked: approx. 100 kg

Weight, packed: 120 kg

Operating temperature range: 15 °C to 35 °C

Relative humidity: 10 to 80 %,

Paraffin wax baths

Number of baths: 3

Volume (liters): 5.0 l per bath

Melt time: approx. 3 hrs

Temperature: 65 °C

Temperature accuracy: + 1 °C

Retort

Capacity: approx. 300 cassettes

Reagent volume : 6.0 Lt.

Temperature (paraffin): 65 °C

Temperature (reagents): ambient or 35 °C

Temperature (cleaning reagents): 65 °C

Temperature accuracy: + 1 °C

Filling time: 5 min

Drain time: 5 min

Impregnation vacuum: -70 kPa (g)

Impregnation pressure: 35 kPa (g)

Fill vacuum: -70 kPa (g)

Drain pressure: 35 kPa (g)



General

Reagent bottles: 9

Cleaning solution bottles: 3

Maximum bottle volume: 4.0 l

Recirculation (pump in/out): ON/OFF

Time between cycles: 15 min

Hardware and Software:

Large color Liquid Crystal Display (LCD)

User-friendly, intelligent software.

Alarm system to reminder wrong operation and troubles

- Multiple specimen protection system.

Capacities:

15 programs that consist of up to 9 reagent and 3 paraffin wax steps.

Time per program step: 0 to 99 hours, 59 minutes.

Up to 300 cassettes can be processed

Wax clean program.

9 reagent bottles.

3 wax baths.

3 cleaning solution bottles.

2 condensate bottle.



Histo-Pro 300

Vacuum Tissue Processor
Extremely Reliable



Distributed by:

Via Brembo, 27 - Milano Italy Tel: +39 0255230061 Fax: +39 0255213764

Histo-Pro 300 Vacuum Tissue Processor



High-resolution LCD display

Easy to operate Control Panel

Drain Paraffin port

14 Container 10 for reagents 3 for cleaning 2 for filtering vapors



Want more Information ? Visit: www.histo-online.com